



ISO:9001:2015 Certified

**Council of Scientific and Industrial Research  
NATIONAL AEROSPACE LABORATORIES**

PB No.1779, HAL Airport Road, Kodihalli, Bangalore – 560017, India

Phone: +91-80-25086040/6041 Fax: +91-80-25269611

Email: purchasek@nal.res.in

**Date: 23-Feb-2021**

**CORRIGENDUM/ADDENDUM**

Tender No. NAL/PUR/MSD/429/20-Z

Tender ID: 2021 CSIR 70416 1 Dated 16-Feb-2021

In continuation of CSIR-National Aerospace Laboratories Tender No. NAL/PUR/MSD/429/20-Z Dated 16-Feb-2021 for "**Procurement of Servo Electric Creep Testing Machine**", kindly note the following **Technical Specification** changes: -

Stroke length is not a mandatory criteria. However, the lower pull rod stroke can be set to 150 mm Max.

This is mainly for static creep testing, therefore the movements of pull rods are expected to return in original position after the completion of test hours.

Other clauses of the bidding document remain unchanged.

  
Sr. Controller of Stores & Purchase  
For and on behalf of CSIR